

## esz AG calibration & metrology

Max- Planck-Str. 16, 82223 Eichenau

has been attested within their quality management to have developed additional requirements to ISO/ IEC 17025:2005 – “*General requirements for the competence of testing an calibration laboratories*” to follow the publication of the American National Standard ANSI/NCSL Z540-1-1994 – “*Calibration Laboratories and Measuring and Test Equipment – General Requirements*” which are not directly addressed within their accreditation.

### Compliance to ANSI/ NCSL Z540-1-1994

Following items (numbered to correspond with the requirements as found in ANSI/NCSL Z540-1-1994, Part I) have been observed and verified as compliant at the laboratory:

- |        |  |   |
|--------|--|---|
| 5.2h)  | The quality manual contains the laboratory’s scope of calibrations   | ✓ |
| 5.4    | The quality system is reviewed at least once a year (management review)  | ✓ |
| 10.2a) | Calibration procedures contain range, tolerance, uncertainty   | ✓ |
|        | The procedures contain the description of the measurement standards  | ✓ |
|        | Procedures are consistent with the accuracy required   | ✓ |
| 11.5   | Tamper resistant seals are affixed to operator accessible controls   | ✓ |
| 13.2   | Each report includes a statement that the certificate shall not be reproduced except in full, without the written approval of the laboratory | ✓ |

This statement of compliance shall apply to all ISO 17025 compliant and ISO 17025 accredited calibrations, to the annex of accreditation and to the scope of accreditation with the accreditation number D-K-15019-01. It is valid for the full period of the accreditation.



---

Marc B. Renz  
Quality Manager



---

Dipl.-Ing. Univ. Philip M. Fleischmann  
Chief Metrologist - CTO